



Type Examination Certificate No CH-MI004-13018-03_EN

Translation

Applicant

Sontex SA
Rue de la Gare 27
2605 Sonceboz-Sombeval
Schweiz

Requirements

Ordinance of 15 February 2006 on Measuring Instruments (SR 941.210) and Ordinance of the FDJP of 19 March 2006 on Heat Meters (SR 941.231)

Directive 2004/22/EC of the European Parliament and of the Council of 31 March 2004 on measuring instruments (MID) and the instrument-specific annex MI-004

Conformity standards:
EN 1434: Edition 2007
OIML R75 : Edition 2002

Type of instrument

Compact heat meter

Type designation

SUPERCAL 739

Accuracy class(es)

3 (according to EN1434)

Characteristics

q_p :	0.6 ... 2.5 m ³ /h
q_p/q_i :	50 ... 100
Nominal pressure:	1.6 MPa
Nominal diameter:	DN15 to DN20

Certificate valid until

30 July 2023

CH-3003 Bern-Wabern, 15 June 2015

Notified body

Conformity Evaluation Body METAS-Cert
No 1259

Gulian Couvreur, Head METAS-Cert

Remark: This certificate is a translation. In case of equivocality, the content from the original certificate is decisive.