



# EU Type Examination Certificate No CH-MI004-13018-05\_EN

## Translation

<i>Applicant</i>	<b>Sontex SA</b> Rue de la Gare 27 2605 Sonceboz-Sombeval Schweiz
<i>Requirements</i>	Ordinance of 15 February 2006 on Measuring Instruments (SR 941.210) and Ordinance of the FDJP of 19 March 2006 on Heat Meters (SR 941.231)  Directive 2014/32/EU of the European Parliament and of the Council of 26 February 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of measuring instruments (MID) and the instrument-specific annex VI (MI-004)
<i>Conformity standards:</i>	EN 1434:2007; EN 1434:2015 OIML R75:2002
<i>Type of instrument</i>	<b>Compact heat meter</b>
<i>Type designation</i>	<b>SUPERCAL 739</b>
<i>Accuracy class(es)</i>	3 (according to EN1434)
<i>Characteristics</i>	$q_p$ : 0.6 ... 2.5 m <sup>3</sup> /h $q_p/q_i$ : 50 ... 100 Nominal pressure: 1.6 MPa Nominal diameter: DN15 to DN20
<i>Certificate valid until</i>	30 July 2023

3003 Berne-Wabern, 30 October 2018

*Approved by* Gulian Couvreur, Head of sector  
METAS-Cert

